

Certificate of Analysis



YSI Incorporated

Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA

Phone 937 767-7241 • 800 765-4974 • Fax 937 767-9353

| | |
|----------------------|---------------------------|
| Product Category | Buffer and solid products |
| Product Model Number | 2363 |
| Lot Number | 25D100662 |
| Method | Verification of mass |

| Determination | Specification | Result |
|---------------|-----------------|--------|
| Average Mass | 1.15-1.35 grams | 1.25 g |

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by: Amanda Fenner

Date: 5/1/2025

