

**FDO<sup>®</sup> 700 IQ H**

**FDO<sup>®</sup> 701 IQ H**

OPTICAL SENSOR FOR DISSOLVED OXYGEN FOR CLASS I DIVISION 2 APPLICATIONS



a xylem brand



For the most recent version of the manual, please visit [www.yisi.com](http://www.yisi.com).

**Contact**

YSI  
1725 Brannum Lane  
Yellow Springs, OH 45387 USA  
Tel: +1 937-767-7241  
800-765-4974  
Email: [environmental@ysi.com](mailto:environmental@ysi.com)  
Internet: [www.yisi.com](http://www.yisi.com)

**Copyright**

© 2014 Xylem Inc.

## FDO® 70x IQ H - Contents

<b>1</b>	<b>Overview</b>	<b>1-1</b>
1.1	How to use this component operating manual	1-1
1.2	Sensor models	1-2
1.3	Recommended fields of application	1-2
1.4	Structure	1-3
<b>2</b>	<b>Safety</b>	<b>2-1</b>
2.1	Safety information	2-1
2.1.1	Hazard warnings in this operating manual	2-1
2.1.2	Safety information on the product	2-1
2.2	Safe operation	2-2
2.2.1	Authorized use	2-2
2.2.2	Requirements for safe operation	2-2
2.2.3	Unauthorized use	2-2
2.3	Hazardous location use	2-3
<b>3</b>	<b>Commissioning</b>	<b>3-1</b>
3.1	Scope of delivery	3-1
3.2	Installation	3-1
3.3	Commissioning / Getting the instrument ready for measuring	3-3
3.4	FDO® 70x IQ H setting table	3-4
<b>4</b>	<b>Measuring / Operation</b>	<b>4-1</b>
4.1	Measuring	4-1
4.2	Function check and user calibration	4-1
4.2.1	General information	4-1
4.2.2	Function check	4-2
4.2.3	User calibration	4-4
4.2.4	Calibration history	4-6
4.2.5	Reactivating previous calibration data	4-7
<b>5</b>	<b>Maintenance, cleaning, replacement parts</b>	<b>5-1</b>
5.1	General maintenance instructions	5-1
5.2	Handling of the sensor cap	5-2
5.3	Exchanging the sensor cap	5-2
5.4	Cleaning the sensor	5-3
5.4.1	Exterior cleaning	5-3
5.4.2	Interior cleaning of sensor cap and sensor head	5-4
5.5	Replacement parts and accessories	5-5
<b>6</b>	<b>What to do if...</b>	<b>6-1</b>
<b>7</b>	<b>Technical data</b>	<b>7-1</b>
7.1	General measurement characteristics	7-1
7.2	Application conditions	7-1

---

7.3	General data .....	7-2
7.4	Instrument safety .....	7-2
7.4.1	General instrument safety .....	7-2
7.4.2	Hazardous location ratings .....	7-3
7.5	Electrical data .....	7-3
7.6	Technical data of the SC-FDO® 700 .....	7-4
7.7	Technical data of the SC-FDO® 701 .....	7-5
<b>8</b>	<b>Contact Information .....</b>	<b>8-1</b>
8.1	Ordering & Technical Support .....	8-1
8.2	Service Information .....	8-1
<b>9</b>	<b>Indexes .....</b>	<b>9-1</b>
9.1	Explanation of the messages .....	9-1
9.1.1	Error messages .....	9-1
9.1.2	Info messages .....	9-3
9.2	Status info .....	9-4

# 1 Overview

## 1.1 How to use this component operating manual

Structure of the IQ SENSORNET operating manual





Fig. 1-1 Structure of the IQ SENSORNET operating manual

The IQ SENSOR NET operating manual has a modular structure like the IQ SENSORNET system itself. It consists of a system operating manual and the operating manuals of all the components used.

Please file this component operating manual into the ring binder of the system operating manual.

### 1.2 Sensor models

The FDO® 70x IQ H consists of a sensor basis equipped with different sensor cap types, depending on the model:



#### Sensor cap

The sensor cap primarily determines the measuring characteristics of the sensor as a whole. Thus the sensor can be adapted to the measuring job by exchanging the type of sensor cap. The specifications of the individual sensor cap types are given in chapter 7 TECHNICAL DATA.

### 1.3 Recommended fields of application

Stationary measurements in water/wastewater applications in Class I Division 2 hazardous locations.

1.4 Structure

Structure




Fig. 1-2 Structure of the D. O. sensor

1	Shaft
2	Plug head connector
3	Fixing ring
4	Sensor membrane
5	Sensor cap with memory chip
6	Gold-plated contact pins for memory chip
7	Measurement window
8	Temperature sensor and locking device
9	Sensor head

**Sensor cap with memory chip**

A memory chip is integrated in the sensor cap. The following data is stored on the memory chip:

- Type designation of the sensor cap
- Serial number
- Data of the factory calibration

**Intelligent membrane (IQMC technology)**

For each membrane the individual calibration values are determined by a factory calibration and stored on the memory chip of the sensor cap, ensuring maximum accuracy over the whole lifetime of the sensor.









## 2 Safety

### 2.1 Safety information

#### 2.1.1 Hazard warnings in this operating manual

The hazard warnings are defined for the following levels of danger:

	<p style="text-align: center;"><b>⚠ DANGER</b></p> <p>DANGER indicates an imminently hazardous situation which, if not avoided, will result in death or serious injury.</p>
	<p style="text-align: center;"><b>⚠ WARNING</b></p> <p>WARNING indicates a potentially hazardous situation which, if not avoided, could result in death or serious injury.</p>
	<p style="text-align: center;"><b>⚠ CAUTION</b></p> <p>CAUTION indicates a potentially hazardous situation which, if not avoided, may result in minor or moderate injury.</p>
	<p style="text-align: center;"><b>NOTICE</b></p> <p>NOTICE is used to address practices not related to personal injury.</p>

#### 2.1.2 Safety information on the product

Note all labels, information signs and safety symbols on the product.

## 2.2 Safe operation

### 2.2.1 Authorized use

The authorized use of the FDO® 70x IQ H comprises its use as a D.O. sensor in the IQ SENSORNET. The technical specifications according to chapter 7 TECHNICAL DATA must be observed. Only operation according to the instructions in this operating manual is authorized.

Any other use is considered to be **unauthorized**. Unauthorized use invalidates any claims with regard to the guarantee.

### 2.2.2 Requirements for safe operation

Note the following points for safe operation:



- The product may only be operated according to the authorized use specified above.
- The product may only be supplied with power by the energy sources specified in this operating manual.
- The product may only be operated under the environmental conditions specified in this operating manual.
- The product or its components may only be opened if this is required for installation and maintenance work and described in the operating manual.

### 2.2.3 Unauthorized use

The product must not be put into operation if:

- it is visibly damaged (e.g. after being transported)
- it was stored under adverse conditions for a lengthy period of time (storing conditions, see chapter 7 TECHNICAL DATA)

2.3 Hazardous location use

	 <b>DANGER</b>
	<p><b>Explosion hazard.</b>                  Only hazardous location rated sensor models ("Hazloc Sensor") must be used in hazardous locations. Read the name plate on the sensor shaft and verify that the sensor is rated for your specific application.</p>



**Rated sensor models**



The following sensor models are rated for hazardous location use:

- FDO® 700 IQ H (YSI-Order No. 207065Y)
- FDO® 701 IQ H (YSI-Order No. 207066Y)

Hazardous location rated sensors are identified by the name plate with the rating details engraved on the sensor shaft:

**Name plate of a Hazloc sensor (example)**

	 Intertek 2001759 CONFORMS TO ##### ##### CERTIFIED TO ##### #####
FDO 700 IQ H  <b>Hazloc Sensor</b> Rated: Class I Division2 Groups ##### Class I Zone 2 Group ##### ##### Control Drawing NO 28123	

	 <b>DANGER</b>
	<p><b>Explosion hazard.</b>                  The associated apparatus MIQ/BB1 (YSI part no. 207001Y) has to be used for connection. Refer to the following documents for proper installation:</p> <ul style="list-style-type: none"> <li>● Control drawing, YSI document no. 28123 (see MIQ/BB1 operating manual)</li> <li>● MIQ/BB1 operating manual</li> <li>● All applicable electrical code regulations.</li> </ul>

	 <b>DANGER</b>
	<p><b>Explosion hazard.</b>                  This product does not meet the requirements of the directive 94/9/EC (ATEX).</p>





### 3 Commissioning

#### 3.1 Scope of delivery

- Sensor, consisting of sensor basis and sensor cap
- Operating manual.

#### 3.2 Installation

**Safety guidelines for installation in a hazardous location**

	 <b>DANGER</b>
	<p><b>Explosion hazard.</b>  <b>Only hazardous location rated sensor models ("Hazloc Sensor") must be used in hazardous locations. Verify the rating on the product name plate on the sensor shaft.</b>  <b>The associated apparatus MIQ/BB1 (YSI part no. 207001Y) has to be used for connection. Refer to the following documents for proper installation:</b></p> <ul style="list-style-type: none"> <li>● <b>Control drawing, YSI document no. 28123 (see MIQ/BB1 operating manual)</b></li> <li>● <b>MIQ/BB1 operating manual</b></li> <li>● <b>All applicable electrical code regulations.</b></li> </ul>

**Connection cable**

A sensor connection cable of the SACIQ type is required to connect the sensor. The cable is available in different lengths.

	 <b>DANGER</b>
	<p><b>Explosion hazard.</b>  <b>For hazardous location use only the cable types listed in the control drawing must be used.</b></p>



How to connect the SACIQ sensor connection cable to the terminal strip of a MIQ module is described in chapter 3 INSTALLATION of the IQ SENSORNET system operating manual.

**Are the plug connections dry?**

Before connecting the sensor and sensor connection cable, please make sure the plug connections are dry. If moisture gets into the plug connections, first dry the plug connections (dab them dry or blow them dry using compressed air).



Do not suspend the sensor on the sensor connection cable. Use a sensor holder or armature. Information on this and other IQ SENSORNET accessories is given in the YSI catalog and on the Internet.

**Connecting the sensor  
to the sensor  
connection cable**

1	Take the protective caps off the plug connections of the sensor and the SACIQ sensor connection cable and keep them safe.
2	Plug the jack of the SACIQ sensor connection cable onto the plug head connector of the sensor. At the same time, rotate the socket so that the pin in the plug head connector (1) clicks into one of the two holes in the socket.
3	Then screw the coupling ring (2) of the sensor connection cable onto the sensor up to the stop.




Fig. 3-1 Connecting the sensor

### 3.3 Commissioning / Getting the instrument ready for measuring

#### Identification in the IQ SENSORNET

Both the sensor cap and sensor basis (sensor without cap) have a series number of their own. The sensor has to be equipped with a sensor cap for a successful login to the IQ SENSORNET. The operable sensor is displayed as follows in the list of sensors:

- *Model:* Model designation of the sensor cap ("SC FDO 70x")
- *Ser. no.:* Series number of the sensor cap
- *Sensor name:* The series number of the sensor basis is preset here. You can change this entry as necessary by entering a user-defined name.

#### Commissioning steps

1	Pull the protective cap off the sensor.
2	If required, assign a user-defined name to the sensor (see relevant IQ SENSORNET system operating manual).
3	When using IQ SENSORNET without automatic air pressure compensation the mean air pressure value or the location altitude has to be entered (for more detailed information see the relevant IQ SENSORNET system operating manual).
4	Set the sensor (see section 3.4).

## 3.4 FDO® 70x IQ H setting table

Setting	Selection/values	Explanation
<i>Measuring mode</i>	<ul style="list-style-type: none"> <li>● <i>Concentration</i></li> <li>● <i>Saturation</i></li> </ul>	Unit of the measured value on the measured value display.
<i>Measuring range Concentration</i>	<ul style="list-style-type: none"> <li>● <i>0 ... 20.00 mg/l</i></li> <li>● <i>0 ... 20.00 ppm</i></li> </ul>	These measuring ranges are available for selection.
<i>Measuring range Saturation</i>	<ul style="list-style-type: none"> <li>● <i>0 ... 200.0 %</i></li> </ul>	The measuring range is set permanently.
<i>Response time t90</i>	<ul style="list-style-type: none"> <li>● 150 .. 300 s (with SC-FDO® 700) or</li> <li>● 60 ... 300 s (with SC-FDO® 701)</li> </ul>	<p>Response time of the signal filter. Depending on the sample matrix, the measured values may fluctuate more or less. A signal filter in the sensor reduces the fluctuation range of the measured value. The signal filter is characterized by the <i>Response time t90</i>. This is the time after which 90 % of a signal change is displayed.</p> <p>The setting range depends on the sensor cap type.</p>
<i>Calibration</i>	<ul style="list-style-type: none"> <li>● <i>Factory calibration</i></li> <li>● <i>User calibration</i></li> <li>● <i>active</i></li> <li>● <i>abort</i></li> </ul>	<p>Determines which calibration data the measured value calculation will be based on. The active calibration is displayed in the calibration history.</p> <p>The selection, <i>User calibration</i> is only displayed if valid data of a <i>User calibration</i> is stored in the sensor.</p> <p><i>active</i> indicates that the sensor is being calibrated.</p> <p>If <i>abort</i> is selected, the active calibration is canceled the next time the setting table is quit with <i>Save and quit</i>.</p>
<i>Test</i>	<ul style="list-style-type: none"> <li>● <i>active</i></li> <li>● <i>abort</i></li> </ul>	<p><i>active</i> indicates that the sensor is being checked.</p> <p>If <i>abort</i> is selected, the active check is canceled the next time the setting table is quit with <i>Save and quit</i>.</p>
<i>Temperature mode</i>	<ul style="list-style-type: none"> <li>● °C</li> <li>● °F</li> </ul>	Unit of the measured temperature value (Celsius, Fahrenheit).



Setting	Selection/values	Explanation
<i>Temp. adjustment</i>	-1.5 K ... +1.5 K	The temperature compensation enables the temperature display to be balanced (shifting of the zero point by $\pm 1.5$ K). Notes: <ul style="list-style-type: none"> <li>● Due to the thermal capacity of the sensor, it is necessary to place it in a container with at least 2 liters of water.</li> <li>● Leave the sensor in this container for at least 15 minutes while stirring occasionally until the balancing can be performed. With temperature differences between the water and sensor <math>&gt; 10</math> °C, leave the sensor for at least 1 hour in this container while stirring occasionally until the balancing can be performed.</li> </ul>
<i>Salinity</i>	<ul style="list-style-type: none"> <li>● <i>On</i></li> <li>● <i>Off</i></li> </ul>	Determines whether the entered salinity should be taken into account.
<i>Salinity input</i>	2.0 ... 70.0	The entry of the salinity enables a salt content correction that compensates for the effect of salt contents $> 0.1$ % on the oxygen measurement. The salt content correction is recommended for measurements in salt-contaminated wastewater (salinity $\geq 2.0$ corresponding to a conductivity of $\geq 3.4$ mS/cm at a reference temperature $T_{REF} = 20$ °C).
<i>Sensor cap data</i>	<ul style="list-style-type: none"> <li>● <i>Do not download</i></li> <li>● <i>Transmit to log book</i></li> </ul>	Generates a log book message with all data stored in the sensor cap when the settings are quit with <i>Save and quit</i> (see section 1.2). When opening the menu again the setting is reset to <i>Do not download</i> .
<i>Save and quit</i>		The system confirms the saving of the settings and the display switches to the next higher level.
<i>Quit</i>		The display switches to the next higher level without saving the new settings.



### Carrying out settings

Using **<S>**, switch from the measured value display to the main menu of the settings. Then navigate to the setting menu (setting table) of the sensor. The exact procedure is given in the relevant IQ SENSORNET system operating manual.



## 4 Measuring / Operation

### 4.1 Measuring

	 <b>WARNING</b>
	<p><b>Chemical or biological hazard.</b>  <b>Contact with the sample can be harmful to the user.</b>  <b>Depending on the type of sample, suitable protective measures must be taken (protective clothing, goggles, etc.).</b></p>

Note the data given in section 7.2 APPLICATION CONDITIONS, especially the minimum immersion depth of the sensor (> 50 mm). The measured value is available immediately on submersing.



If keeping the sensor clean is a problem, we recommend using the compressed air-driven cleaning system with the CH cleaning head (see section 5.5 REPLACEMENT PARTS AND ACCESSORIES).

### 4.2 Function check and user calibration

#### 4.2.1 General information

#### Factory calibration

The FDO® 70x IQ H is factory calibrated. In the recommended application (see section 1.3 RECOMMENDED FIELDS OF APPLICATION), the measuring characteristics of the sensor cap remain stable for the specified service life. Thus, a user calibration is not usually required.

#### When does a function check or user calibration make sense?


A function check or user calibration can be useful in the following special cases:

- If the measured values appear to be implausible and it is assumed that the service life of the sensor cap is over
- Routinely within the framework of the company quality assurance.

#### Check or calibration medium

Select one of the following two variants depending on the air temperature at the calibration site:

- With air temperatures over 5 °C, the function check and user calibration ideally take place in water vapor-saturated air. To do so, position the sensor approx. 2 cm above a water surface, for example in a narrow bucket or similar container with water. The membrane must be clean and dry for this.



- With air temperatures under 5 °C we recommend performing the function check and user calibration not in air but in air-saturated water that has a higher temperature. You obtain air-saturated water by pouring water several times in and out of two vessels so that it sparkles.

### 4.2.2 Function check

A function check is the simplest way to determine whether the sensor needs to be cleaned or user-calibrated.

#### Principle

The function check can either be done in water vapor-saturated air or in air-saturated water (see CHECK OR CALIBRATION MEDIUM on page 4-1).

#### Steps

Generally, a check on the IQ SENSORNET is carried out as follows. System specific details are given in the respective IQ SENSORNET system operating manual.

1	Switch to the measured value display with <b>&lt;M&gt;</b> and select the FDO® 70x IQ H sensor.
2	Press <b>&lt;C&gt;</b> . The next step switches on the maintenance condition for the sensor. A corresponding message appears on the display.
3	Confirm the note with <b>&lt;OK&gt;</b> . The maintenance condition is active.
4	Select the <i>TEST</i> procedure and press <b>&lt;OK&gt;</b> .
5	Put the sensor into the calibration position (water vapor-saturated air or air-saturated water - see section 4.2.1).
6	Press <b>&lt;OK&gt;</b> . The sensor starts the check. The display switches to the measured value display. The <i>CAL</i> indicator flashes instead of the main measured value. At the same time, the momentary relative slope flashes as the secondary measured value with the addition of <i>TEST</i> . The process ends automatically as soon as the measured values meet the criterion for the stability control. With a great temperature difference between the sensor and the environment this may take a while. Subsequently, the main measured value and temperature are displayed.

7	Put the sensor in the measuring position again.
8	Wait for the measured value to be largely stable (temperature adjustment).
9	Switch off the maintenance condition.

### Canceling the check

As long as the determination of the relative slope has not yet been started (step 6), you can quit the check with **<M>** or **<ESC>**.

The running determination of the relative slope (after pressing **<OK>** in step 6) can be aborted as follows:

1	Open the setting table (see section 3.4).
2	In the <i>Test</i> menu item, select the <i>abort</i> setting and then quit the setting table with <i>Save and quit</i> .

### Evaluation

The result of the check is entered in the log book of the sensor. If the relative slope is outside the tolerance range (0.90 ... 1.10), the check is assessed as erroneous.



The information recorded in the log book entry is partly identical with that of the calibration history. The calibration history is described in detail in section 4.2.4 CALIBRATION HISTORY.

### 4.2.3 User calibration

#### Principle

The user calibration can either be done in water vapor-saturated air or in air-saturated water (see CHECK OR CALIBRATION MEDIUM on page 4-1). With the calibration procedure, the relative slope of the sensor is determined. The calibration is evaluated based on the relative slope and the intensity (successful <-> unsuccessful).

The result of the user calibration is stored in the calibration record and calibration history respectively and can be viewed afterwards (see respective IQ SENSORNET system operating manual).



We want to point out that the factory calibration of the membrane is highly precise due to the IQMC technology. If you still want to carry out a user calibration, influences due to environmental conditions have to be taken into account.

#### Steps

Generally, a user calibration on the IQ SENSORNET is carried out as follows. System specific details are given in the respective IQ SENSORNET system operating manual.

1	Switch to the measured value display with <b>&lt;M&gt;</b> and select the FDO® 70x IQ H sensor.
2	Press <b>&lt;C&gt;</b> . The next step switches on the maintenance condition for the sensor. A corresponding message appears on the display.
3	Confirm the note with <b>&lt;OK&gt;</b> . The maintenance condition is active.
4	Select the <i>CALIBRATION</i> procedure and press <b>&lt;OK&gt;</b> .
5	Put the sensor into the calibration position (water vapor-saturated air or air-saturated water - see section 4.2.1).
6	Press <b>&lt;OK&gt;</b> . The sensor determines the calibration data. The display switches to the measured value display. The <i>CAL</i> indicator flashes instead of the main measured value. At the same time, the momentary relative slope flashes as the secondary measured value. The process ends automatically as soon as the measured values meet the criterion for the stability control. With a great temperature difference between the sensor and the environment this may take a while. Subsequently, the main measured value and temperature are displayed.
7	If the user calibration was successful, bring the sensor into the measuring position again.
8	Wait for the measured value to be largely stable (temperature adjustment).
9	Switch off the maintenance condition.

**Canceling the user calibration**

As long as the determination of the calibration data has not yet been started (step 6), you can quit the calibration routine with **<M>** or **<ESC>**.

The running determination of calibration data (after pressing **<OK>** in step 6) can be aborted as follows:

1	Open the setting table (see section 3.4).
2	In the <i>Calibration</i> menu item, select the <i>abort</i> setting and then quit the setting table with <i>Save and quit</i> .

After the user calibration was canceled the sensor again works with the calibration data that was used before the canceled user calibration.

**Possible results of the user calibration**

The calibration data is evaluated by the system. A calibration can have the following results:

Display	Explanation
Measured value display	Sensor was successfully calibrated. The calibration data can be viewed in the calibration history (section 4.2.4).
"----"	The sensor could not be calibrated. The sensor is blocked for further measurement. Notes on possible causes are given in the log book of the sensor.

### 4.2.4 Calibration history

Calibration history  
(available in the  
IQ SENSORNET system  
2020 XT only)

MIQ/T2020		15 Dec 2006 00:13		[Icons]	
Calibration history of selected sensor 330					
S01 SC FDO 700		03270001			
Date	Rel.slope	Intens.	Res		
-----					
Factory	1.00	+	o.k.		
-----					
15.12.2006	0.77	+	Error		
04.11.2006	0.96	+	o.k.		
Range of tolerance 0.90 .. 1.10					
Return ESC					

Currently active calibration

Chronological list of the last user calibrations

Fig. 4-1 Calibration history FDO® 70x IQ H

The calibration history contains the following information:

<b>Date</b>	Calibration date ( <i>Factory</i> = factory calibration)
<b>Rel.slope</b>	Relative slope (non-dimensional)
<b>Intens.</b>	Intensity: + : sufficient - : too low
<b>Res</b>	Evaluation of the calibration o.k. : Calibration successful. The new calibration data was taken over for measurement. Error : Calibration unsuccessful. Invalid calibration data were discarded. Measurement was continued with the last valid values.



The calibration history is stored in the sensor basis. If another sensor cap (with different series number) is mounted, the calibration history is deleted and cannot be restored afterwards.



#### 4.2.5 Reactivating previous calibration data

The FDO® 70x IQ H enables you to reactivate the last valid user calibration or the factory calibration. Thus you can immediately go on measuring if a calibration procedure failed or you suspect that the calibration conditions were not optimally met.



Reactivating old calibration data is only a temporary solution. Please take into account that the sensor might thus possibly provide incorrect measured values. Please make sure the sensor works correctly by carrying out a function check and/or recalibration.


#### Reactivating calibration data

1	Open the setting table (see section 3.4).
2	In the <i>Calibration</i> menu item, select the <i>User calibration</i> or <i>Factory calibration</i> setting and then quit the setting table with <i>Save and quit</i> .



## 5 Maintenance, cleaning, replacement parts

### 5.1 General maintenance instructions

	<b>⚠ WARNING</b>
	<b>Chemical or biological hazard.</b> <b>Contact with the sample can be harmful to the user.</b> <b>Depending on the type of sample, suitable protective measures must be taken (protective clothing, goggles, etc.).</b>

#### Maintenance condition

We recommend to switch on the maintenance condition each time the sensor is taken out of the measuring position. This avoids unwanted reactions of linked outputs. For more detailed information on the maintenance condition please refer to the respective IQ SENSORNET system operating manual.



We do not recommend unscrewing the sensor from the sensor connection cable in order to do maintenance work. Otherwise, moisture and/or dirt can get into the plug connection where it can cause contact problems.


If you would like to disconnect the sensor from the sensor connection cable, please note the following points:

- Before disconnecting the sensor from the SACIQ sensor connection cable, the sensor has to be cleaned on the outside (see section 5.4.1).
- Unscrew the sensor from the SACIQ sensor connection cable.
- Place a protective cap on the sensor plug head connector and on the SACIQ sensor connection cable each time so that no moisture or dirt can get on the contacting surfaces.
- In corrosive environments, close the socket of the sensor connection cable with the screwable SACIQ-Plug when it is dry in order to protect the electrical contacts from corrosion. The protective plug is available as an accessory (see section 5.5 REPLACEMENT PARTS AND ACCESSORIES).

### 5.2 Handling of the sensor cap


Despite its exterior robustness, the sensor is an optical high precision instrument. Therefore, special care should be taken when doing any maintenance or cleaning work:

- Dirt and moisture under the sensor cap can affect the functioning and shorten the service life of the sensor cap. Therefore, make sure the working environment is clean and dry prior to removing the sensor cap.
- Please do not touch the outer sensor membrane with your fingers. Touch the sensor cap at the sides only (shaded area in figure on the left).
- Avoid any great mechanical stress of the sensor membrane (pressure, scratches).
- Exposure to light, particularly daylight of the interior of the sensor cap will, by-and-by, affect the measurement characteristics and shorten the service life of the sensor cap. Therefore, the interior of the sensor cap should not be exposed to direct sunlight. Avoid any exposure to light that exceeds the extent required for necessary maintenance and cleaning activities. Store dismantled sensor caps in a light-protected environment only.



### 5.3 Exchanging the sensor cap


Removing the sensor cap



1	Pull the sensor out of the sample.
2	Clean the outside of the sensor (see section 5.4.1).
3	Unscrew the fixing ring from the sensor by hand.
4	Thoroughly clean and dry the sensor head once again.
5	Grasp the sensor cap on the sides (arrows in figure on the left) and remove it by pulling it away from the sensor in a <u>straight</u> upward direction.

	<b>NOTICE</b>
<p><b>Do not push any tools or other sharp objects between the sealing surfaces. This might damage the sealing surfaces.</b></p>	

### Mounting the sensor cap



6	Check the front surface of the sensor for absolute cleanness and clean it if necessary (see section 5.4.1).
7	Thoroughly clean the thread of the fixing ring.
8	Place the new sensor cap on the sensor so that the temperature sensor fits into the hole inside the sensor cap (see figure opposite).
9	Put the fixing ring on the sensor head and screw it tight by hand.

## 5.4 Cleaning the sensor

### 5.4.1 Exterior cleaning


Dirt on the sensor can affect the measuring characteristics. Biological deposits for example, consume oxygen and can, when occurring on the sensor cap membrane, impair the responding behavior and cause values that are too low. Therefore, we recommend regular visual inspections and exterior cleaning as necessary.

Pay attention to the following points for cleaning:

- First, thoroughly rinse the sensor with tapwater to remove loosely adhering dirt.
- Rough dirt on the sensor shaft can be brushed off with a soft brush. **Attention:** Do not use the brush in the area of the sensor membrane. Risk of damage!
- The sensor cap including the sensor membrane should be wiped with a soft and moist microfiber cloth.
- In the case of persisting dirt you can add some household washing-up liquid to the tapwater. **Attention:** Never use any alcohol for cleaning!
- If there are any salt or lime deposits, the sensor can be cleaned with aqueous citric acid solution (10 % weight percent).

### 5.4.2 Interior cleaning of sensor cap and sensor head

If moisture or dirt have penetrated under the sensor cap, e.g. because the sensor cap is damaged, you can make the sensor ready for operation again as follows:

	<b>NOTICE</b>
	<b>Only use <u>nonabrasive, alcohol-free</u> detergents, as otherwise the optical surfaces could be damaged.</b>

1	Remove the sensor cap (see section 5.3).
2	Clean the sensor head and sensor cap: <ul style="list-style-type: none"><li>– Rinse all inner surfaces with tapwater</li><li>– Remove contamination containing fat and oil with warm water and household washing-up liquid</li></ul>
3	Dry all surfaces with a clean, lint free cloth. A lint-free cloth such as a microfiber cloth used to clean eyeglasses is suitable.
4	Allow the sensor and sensor cap to dry completely at a dry location so moisture can evaporate even from corners difficult to access. When doing, so protect the inside of the sensor cap from light.



If the sensor cap is visibly damaged it has to be replaced.

### 5.5 Replacement parts and accessories

Description	Model	Order no.
Sensor cap	SC-FDO® 700	201 654Y
Sensor cap	SC-FDO® 701	201 655Y
Protective screw cap for sensor connection cable	SACIQ-Plug	480 065Y

#### Components for compressed air cleaning system

Description	Model	Order no.
Cleaning head	CH	900 107Y
Active valve module (does not require a free relay output in the IQ SENSORNET system)	MIQ/CHV PLUS	480 018Y
Valve module for the IQ SENSORNET system 182 (external relay and compressed air supply)	DIQ/CHV	472 007Y
Air compressor to be triggered by relay, power supply 115 V AC	Cleaning Air Box - 115 VAC	480 017Y
Air compressor to be triggered by relay, power supply 230 V AC	Cleaning Air Box - 230 VAC	480 019Y



Information on further IQ SENSORNET accessories is given in the YSI catalog and on the Internet.





## 6 What to do if...

**Sensor does not appear in the measured value display and list of sensors**

Cause	Remedy
<ul style="list-style-type: none"> <li>– Sensor cap not mounted or defective</li> </ul>	<ul style="list-style-type: none"> <li>– Sensor cap (see section 5.3)</li> <li>– Replace a defective sensor cap as necessary</li> </ul>

**Measured value implausible**

Cause	Remedy
<ul style="list-style-type: none"> <li>– Coating on sensor cap</li> </ul>	<ul style="list-style-type: none"> <li>– Clean the outside of the sensor (see section 5.4.1)</li> </ul>
<ul style="list-style-type: none"> <li>– Service life of the sensor cap over</li> </ul>	<ul style="list-style-type: none"> <li>– Check the sensor (see section 4.2.2)</li> <li>– Replace the sensor cap as necessary (see section 5.3)</li> </ul>
<ul style="list-style-type: none"> <li>– Dirt inside the sensor cap and sensor head.</li> <li>– Fixing ring not properly tightened.</li> <li>– Sensor cap untight or defective.</li> </ul>	<ul style="list-style-type: none"> <li>– Dismantle sensor cap</li> <li>– Clean the insides of the sensor cap and sensor head (see section 5.4.2)</li> <li>– Mount the sensor cap correctly and tighten the fixing ring as far as it will go (see section 5.3)</li> <li>– Replace a defective sensor cap as necessary</li> </ul>

**Measured value fluctuating heavily**

Cause	Remedy
<ul style="list-style-type: none"> <li>– Coating on sensor cap</li> </ul>	<ul style="list-style-type: none"> <li>– Clean the outside of the sensor (see section 5.4.1)</li> </ul>
<ul style="list-style-type: none"> <li>– Service life of the sensor cap over</li> </ul>	<ul style="list-style-type: none"> <li>– See point "Measured value implausible"</li> </ul>
<ul style="list-style-type: none"> <li>– Dirt inside the sensor cap and sensor head.</li> <li>– Fixing ring not properly tightened.</li> <li>– Sensor cap untight or defective.</li> </ul>	<ul style="list-style-type: none"> <li>– See point "Measured value implausible"</li> </ul>

<b>Measured value invalid ( "----" displayed)</b>	<b>Cause</b>	<b>Remedy</b>
	<ul style="list-style-type: none"><li>– User calibration unsuccessful. The sensor is blocked for measurement.</li></ul>	<ul style="list-style-type: none"><li>– As a temporary measure to quickly restore the readiness for service: Activate the factory calibration (see section 4.2.5)</li><li>– For exact measurements, carry out a function check and repeat the user calibration if necessary.</li></ul>
<b>Incorrect temperature display</b>	<b>Cause</b>	<b>Remedy</b>
	<ul style="list-style-type: none"><li>– Temperature sensor defective</li></ul>	<ul style="list-style-type: none"><li>– Return the sensor</li></ul>

## 7 Technical data

### 7.1 General measurement characteristics



The measurement characteristics are primarily determined by the sensor cap type. The relevant data are given in the sections 7.6 and 7.7.

**Measuring principle**  
**Measurement in water**  
**Measurement in salt-contaminated wastewater**

Optical measurement based on photoluminescence.

According to solubility function according to ISO 5814

Salinity input from 2.0 ... 70.0; corresponds to 3.4 mS/cm ... 86.2 mS/cm at T<sub>REF</sub> 20 °C (salinity measurement according to IOT = International Oceanographic Tables)

**Temperature measurement**

Temperature sensor	Integrated NTC
Measuring range	- 5 °C ... + 50 °C (23 ... 122 °F)
Accuracy	± 0.5 K
Resolution	0.1 K

**Temperature compensation**

In the range -5 °C ... + 50 °C (23 ... 122 °F)

### 7.2 Application conditions

**Pressure resistance**

Sensor with connected SACIQ sensor connection cable:

Max. allowed overpressure	10 <sup>6</sup> Pa (10 bar)
---------------------------	-----------------------------

The sensor meets all requirements according to article 3(3) of 97/23/EC ("pressure equipment directive").

**Type of protection**

Sensor with connected SACIQ sensor connection cable:

IP 68, 10 bar (10<sup>6</sup> Pa)

**Immersion depth**


Min. 10 cm; max. 100 m depth

**Operating position**

Any

**7.3 General data**

**Dimensions**



**Weight (without sensor connection cable)**

Approx. 900 g

**Connection technique**

Connection via SACIQ sensor connection cable

**Material**

Shaft	V4A stainless steel 1.4571 *
Plug head connector housing	POM
Sensor head	POM and PVC
Sensor cap	See section 7.6 or section 7.7
Plug, 3-pole	ETFE (blue) Tefzel®
Fixing ring	POM

\* Stainless steel can be corrodible if there are chloride concentrations of 500 mg/L or more.

**Automatic sensor monitoring (SensCheck function)**

Monitoring of the membrane function

**7.4 Instrument safety**

**7.4.1 General instrument safety**

Applicable norms

- UL 61010-1
- CAN/CSA C22.2 # 61010.1

### 7.4.2 Hazardous location ratings

#### Applicable directives and standards

In addition to the standards listed in chapter 7.4.1, the FDO<sup>®</sup> 70x IQ H conforms to the following directives and standards:

- ANSI/ISA-12.12.01
- CAN/CSA C22.2 # 213

The FDO<sup>®</sup> 70x IQ H is a non-incendive electrical equipment for use in Class I Division 2 hazardous locations.

#### Hazardous location rating

FDO <sup>®</sup> 700 IQ H FDO <sup>®</sup> 701 IQ H	Class I Division 2 Groups A, B, C, D T6 Class I Zone 2 Group IIC T6
Associated apparatus	MIQ/BB1, YSI part no. 207001Y
Control Drawing	YSI document no. 28123

### 7.5 Electrical data

Voltage range	14 V DC ... 24 V DC via the MIQ/BB1 barrier box (=associated apparatus)
Power consumption	0.7 W
Protective class	III

### 7.6 Technical data of the SC-FDO® 700

**Measuring ranges and resolutions**

D. O. partial pressure 0 ... 400 hPa

Measuring mode	Adjustable measuring range	Resolution
D. O. concentration	0 ... 20.00 mg/L 0 ... 20.00 ppm	0.01 mg/L 0.01 ppm
D. O. saturation	0 ... 200.0 %	0.1 %

**Measurement accuracy**

In the range	Accuracy
< 1 mg/L (ppm)	± 0.05 mg/L (ppm)
> 1 mg/L (ppm)	± 0.1 mg/L (ppm)

**Repeatability**

± 0.05 mg/L (ppm) according to EN ISO 15839

**Response time**

t<sub>90</sub> (90 % of the final display value after) < 150 s according to EN ISO 15839  
 t<sub>95</sub> (95 % of the final display value after) < 200 s according to EN ISO 15839

**Minimum approach flow**

No incident flow required

**Interferences**

Non-sensitive to against hydrogen sulfide, chlorine, and ionogenic substances

**Allowed temperature range**

Measuring medium	- 5 °C ... + 50 °C (23 ... 122 °F)
Storage/transport	- 10 °C ... + 50 °C (14 ... 122 °F)

**Allowed pH range of the test sample**

4 ... 12

**Calibration procedure**

Factory calibration by means of IQMC (Intelligent Membrane Calibration) procedure. User calibration possible in water vapor-saturated air or air-saturated water.

**Material**

PMMA, PVC and silicone

**Working life**

2 years guaranteed with authorized use

## 7.7 Technical data of the SC-FDO<sup>®</sup> 701

### Measuring ranges and resolutions

D. O. partial pressure 0 ... 400 hPa

Measuring mode	Adjustable measuring range	Resolution
D. O. concentration	0 ... 20.00 mg/L 0 ... 20.00 ppm	0.01 mg/L 0.01 ppm
D. O. saturation	0 ... 200.0 %	0.1 %

### Measurement accuracy

In the range	Accuracy
< 1 mg/L (ppm)	± 0.05 mg/L (ppm)
> 1 mg/L (ppm)	± 0.1 mg/L (ppm)

### Repeatability

± 0.05 mg/L (ppm) according to EN ISO 15839

### Response time

t<sub>90</sub> (90 % of the final display value after) < 60 s according to EN ISO 15839

t<sub>95</sub> (95 % of the final display value after) < 80 s according to EN ISO 15839

### Minimum approach flow

No incident flow required

### Interferences

Non-sensitive to against hydrogen sulfide, chlorine, and ionogenic substances

### Allowed temperature range

Measuring medium	-5 °C ... + 40 °C (23 ... 104 °F)
Storage/transport	- 10 °C ... + 40 °C (-33 ... 104 °F)

### Allowed pH range of the test sample

4 ... 12

### Calibration procedure

Factory calibration by means of IQMC (Intelligent Membrane Calibration) procedure. User calibration possible in water vapor-saturated air or air-saturated water.

### Material

PMMA, PVC and silicone

### Working life

6 months guaranteed for authorized use under the specified environmental conditions





## 8 Contact Information

### 8.1 Ordering & Technical Support

Telephone: (800) 897-4151  
(937) 767-7241  
Monday through Friday, 8:00 AM to 5:00 PM ET

Fax: (937) 767-1058

Email: [environmental@ysi.com](mailto:environmental@ysi.com)

Mail: YSI Incorporated  
1725 Brannum Lane  
Yellow Springs, OH 45387  
USA

Internet: [www.ysi.com](http://www.ysi.com)

When placing an order please have the following information available:

YSI account number (if available)	Name and Phone Number
Model number or brief description	Billing and shipping address
Quantity	Purchase Order or Credit Card

### 8.2 Service Information

YSI has authorized service centers throughout the United States and Internationally. For the nearest service center information, please visit [www.ysi.com](http://www.ysi.com) and click 'Support' or contact YSI Technical Support directly at 800-897-4151.

When returning a product for service, include the Product Return form with cleaning certification. The form must be completely filled out for an YSI Service Center to accept the instrument for service. The Product Return form may be downloaded at [www.ysi.com](http://www.ysi.com) and clicking on the 'Support' tab.



## 9 Indexes

### 9.1 Explanation of the messages

This chapter contains a list of all the message codes and related message texts that can occur in the log book of the IQ SENSORNET system for the FDO<sup>®</sup> 70x IQ H sensor.



Information on the contents and structure of the log book, and how to call it up, is given in the LOG BOOK chapter of the IQ SENSORNET system operating manual.



The last three digits of the message code identify the source of the message:

- 334 = SC FDO 700 (sensor cap SC-FDO<sup>®</sup> 700)
- 335 = SC FDO 701 (sensor cap SC-FDO<sup>®</sup> 701)
- 531 = FDO700IQ (sensor basis / component class, adapter ADA)

#### 9.1.1 Error messages

<b>Message code</b>	<b>Message text</b>
EA133x	<i>Meas. range exceeded or undercut</i> * <i>Check process</i> * <i>Select other meas. range</i>
EA233x	<i>Sensor temperature too high!</i> * <i>Check process and application</i>
EA333x	<i>Sensor temperature too low!</i> * <i>Check process and application</i>
EAP33x	<i>Measurement interfered</i> * <i>SensCheck: Sensor cap is missing, leaky, depleted, or defective</i> * <i>Clean sensor and space between cap and sensor according to op.instructions</i> * <i>Screw on sensor cap securely</i> * <i>Replace sensor cap</i>
EAP531	<i>Measurement interfered</i> * <i>SensCheck: Sensor cap is missing, leaky, depleted, or defective</i> * <i>Clean sensor and space between cap and sensor according to op.instructions</i> * <i>Screw on sensor cap securely</i> * <i>Replace sensor cap</i>

<b>Message code</b>	<b>Message text</b>
EC833x	<i>Sensor could not be calibrated, sensor blocked for measurement Cause: instable signal * Check temperature adjustment * Check calibration conditions (see operating manual) * Repeat calibration</i>
EC933x	<i>Calibration error, measurement disabled Cause: Sensor cap is missing, leaky, depleted, or defective * Clean sensor and space between cap and sensor according to op.instructions * Screw on sensor cap securely * Repeat calibration * Replace sensor cap</i>
EI133x	<i>Operational voltage too low * Check installation and cable lengths, Follow installation instructions * Power supply module(s) overloaded, add power supply module(s) * Check terminal and module connections * Defective components, replace components</i>
EI233x	<i>Operational voltage too low, no operation possible * Check installation and cable lengths, Follow installation instructions * Power supply module(s) overloaded, add power supply module(s) * Check terminal and module connections * Defective components, replace components</i>
ES133x	<i>Component hardware defective * Contact service</i>


### 9.1.2 Info messages

<b>Message code</b>	<b>Message text</b>
IC133x	<i>Sensor has been successfully calibrated * For calibration data, see calibration history</i>
IC333x	<i>Factory calibration has been activated. Make sure the sensor operates correctly.</i>
IC433x	<i>Last valid calibration has been activated. Make sure the sensor operates correctly.</i>
IC533x	<i>Invalid user calibration has been replaced by last valid user calibration. Caution! Wrong measured values possible. Carry out a new successful calibration to make sure the sensor operates correctly.</i>
IC633x	<i>The invalid user calibration was replaced by the factory calibration. Caution! Wrong measured values possible. Check whether the sensor operates correctly or carry out a new successful calibration.</i>
IC733x	This message displays the result of the last check (see section 4.2.2)
IS133x	This message displays the information stored in the sensor cap (see section 1.2)

## 9.2 Status info

The status info is a coded piece of information on the current status of a sensor. Each sensor sends this status info to the controller. The status info of sensors consists of 32 bits, each of which can have the value 0 or 1.

### Status info, general structure



The bits 0 - 15 are reserved for general information.  
 The bits 16 - 21 are reserved for internal service information.

You obtain the status info:

- via a manual query in the *Einstellungen/Settings/Service/List of all components* menu (see system operating manual)
- by an automated query
  - from a superordinate process control (e. g. when connected to the Profibus)
  - from the IQ Data Server (see IQ SENSORNET Software Pack operating manual)



The evaluation of the status info, e.g. in the case of an automated query, has to be made individually for each bit.

### Status info

Status bit	Explanation
<b>Bit 0</b>	<i>Component hardware defective</i>
<b>Bit 1</b>	Sensor cap missing, untight, depleted or defective
<b>Bit 2-31</b>	-



# Xylem |'zīləm|

- 1) The tissue in plants that brings water upward from the roots;
- 2) a leading global water technology company.

We're 12,500 people unified in a common purpose: creating innovative solutions to meet our world's water needs. Developing new technologies that will improve the way water is used, conserved, and re-used in the future is central to our work. We move, treat, analyze, and return water to the environment, and we help people use water efficiently, in their homes, buildings, factories and farms. In more than 150 countries, we have strong, long-standing relationships with customers who know us for our powerful combination of leading product brands and applications expertise, backed by a legacy of innovation.

**For more information on how Xylem can help you, go to [www.xylem.com](http://www.xylem.com)**



a **xylem** brand

YSI  
1725 Brannum Lane  
Yellow Springs, OH 45387  
Tel: +1 937-767-7241; 800-765-4974  
Fax: +1 937-767-1058  
Email: [environmental@ysi.com](mailto:environmental@ysi.com)  
Web: [www.ysi.com](http://www.ysi.com)

©Xylem Inc